



NOBAMED Paul Danz AG

RUDADETEC® FINGERTIP plaster

REF 040100

Product Description, Intended use, Application

Blue plaster strips for fingertips with worked-in aluminum strips which render the plasters detectable – specially produced for the food industry. For covering of smaller, primary healing wounds.

Composition

Polyethylene, polyacrylate adhesive, aluminum

Contraindications

The product should not be used in the case of a known allergy against the material.

Incident reporting

According to MDR (EU) 2017/745, if serious incidents occur in relation to the device, they must be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.

Normative and Regulative Requirements, Common Standards

Medical Device according to MDD 93/42/EEC, MDR (EU) 2017/745.

The product does not contain dangerous toxic substances according to REACH.

Sterilization of the product complies with DIN EN 11135.

Packaging

Primary:	paper packaging
Secondary:	folding box made of cellulose
Tertiary:	carton made of cellulose

Symbols used in labelling

Explanation at www.nobamed.com



Marking on all packaging levels with CE and according to DIN EN ISO 15223-1- and DIN EN ISO 20417

Storage and Transport

Dry and dustfree

Sterile device

Before using a sterile product, visually inspect the packaging to ensure that it is intact.

Single use device

Reusing a single use medical device can lead to microbiological danger. Reprocessing for reuse can decrease the product's performance significantly.

Disposal

According to locally applicable legal regulations and standards of infection prophylaxis.