



## RUDADETEC®

REF 040212

### Product Description, Intended use, Application

Blue plaster strips (2.0 cm x 12.0 cm) with worked-in aluminum strips which render the plasters detectable – specially produced for the food industry. For covering of smaller, primary healing wounds.

### Composition

Polyethylene, polyacrylate adhesive, aluminum

### Contraindications

The product should not be used in the case of a known allergy against the material.

### Incident reporting

According to MDR (EU) 2017/745, if serious incidents occur in relation to the device, they must be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.

### Normative and Regulative Requirements, Common Standards

Medical Device according to *MDD 93/42/EEC, MDR (EU) 2017/745*.

Sterilization of the product complies with DIN EN 11135.

The product does not contain dangerous toxic substances according to REACH.

### Packaging

Primary:	paper packaging
Secondary:	folding box made of cellulose
Tertiary:	carton made of cellulose

### Symbols used in labelling

Explanation at [www.nobamed.com](http://www.nobamed.com)



Marking on all packaging levels with CE and according to DIN EN ISO 15223-1- and DIN EN ISO 20417

### Storage and Transport

Dry and dustfree

### Sterile device

Before using a sterile product, visually inspect the packaging to ensure that it is intact.

### Single use device

Reusing a single use medical device can lead to microbiological danger. Reprocessing for reuse can decrease the product's performance significantly.

### Disposal

According to locally applicable legal regulations and standards of infection prophylaxis.