

NOBAMED Paul Danz AG

NOBASTRETCH®-fein

REF 480710

Product Description, Intended use, Application

The brown, fine woven, long-stretch compression bandage (10 cm x 7 m) has a stretchability of approx. 170 % and is permanently elastic. Long-stretch bandages are mainly used to create bandages that will support and relieve muscles or joints in cases of disorders, for example in the case of sports injuries. They are also used in cases of phlebologic indications like superficial varicose veins and for the follow-up treatment of venous conditions. Long-stretch bandages may also be used to complement compression bandages during the day to flush out oedema. If so used, and if used as phlebologic compression bandage in general, the relevant medical guidelines must be observed. Due to the high resting pressure NOBASTRETCH®fein is not left applied overnight or in the case of longer periods of rest during the day.

Composition

Cotton, polyamide, polyurethane

Contra-Indications

The product should not be used in the case of a known allergy against the material.

Do not use in the case of deep venous disorders. In the case of a medical compression therapy the contra-indications and risks mentioned in the relevant medical guidelines must be observed.

Incident reporting

According to MDR (EU) 2017/745, if serious incidents occur in relation to the device, they must be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.

Normative and Regulative

Requirements, Common Standards Medical Device according to *MDD 93/42/EEC, MDR (EU) 2017/745.*

The product does not contain dangerous toxic substances according to REACH.

Packaging

Primary:

folding box made of cellulose

Secondary:

carton made of cellulose

Symbols used in labelling Explanation at www.nobamed.com



Marking on all packaging levels with CE and according to DIN EN ISO 15223-1- and DIN EN ISO 20417

Storage and transport

To be stored in a dry and dust-free environment

Single use device

Reusing a single use medical device can lead to microbiological danger. Reprocessing for reuse can decrease the product's performance significantly.

Disposal

According to locally applicable legal regulations and standards of infection prophylaxis.