

X-RAY MULLKOMPRESSEN

REF 612520 (as of LOT 36014)

Contents of this procedure pack:

5 x gauze swab, X-Ray 10 x 20 cm, 12 ply, NOBAMED Paul Danz AG, MD C € 0123

1 x wrapping material 40 x 40 cm, VP Feuchtwangen, MD ←

Product Description, Intended use, Application

The sterile procedure pack consists of 5 x-ray-gauze swabs and a wrapping material which can be used as sterile underlay.

The disposable 17-thread gauze swabs with raw edges folded back are used for wound care in the OR area, for wound cleansing and for the absorption of blood and wound fluid.

Composition

Cotton, x-ray thread made of barium sulphate, PVC, DOTP

Contra-Indications

No contra-indications known. The product should not be used in the case of a known allergy against the material.

This product is for single and temporary use on one patient and must be applied by specially trained medical personnel. Any information for use that is available for individual products must be observed!

Normative and Legal Requirements

Procedure Pack according to Art. 22, Medical Device Regulation (EU) 2017/745, risk class IIa.

The compatibility of the customized medical devices or other products has been assessed with due regard for the indications of the respective manufacturers, and operations have been performed in accordance with the aforementioned indications.

The manufacturers' intended use is not changed.

The systems and procedure packs are packed in compliance with a validated procedure.

All activity is properly monitored and controlled as part of our QM system, in accordance with DIN EN ISO 13485:2016.

A validated sterilization process is used for sterile procedure packs, taking into account the manufacturers' instructions, so that product functionality and product safety of the components are maintained after a system or procedure pack has been compiled and sterilized.

Sterilization and packaging are subject to supervision of our Notified Body TÜV Süd PS 0123 pursuant to Annex XI, Part A.

The gauze complies with the requirements of DIN EN 14079 type 17. The product does not contain dangerous toxic substances according to REACH. It has CE marking and DIN EN ISO 15223-1 labels on all its packaging

Storage and transport

I in a dry and dust-free environment, protect from sunlight.

Symbols used in labelling

Explanation at www.nobamed.com











Product Data Sheet: December 20, 2022 [Rev 9], replaces: February 24, 2022 [Rev 8]