



MULLTUPFER-steril

packed in perforated 2-chamber packages
2+2 gauze balls

REF 733220

Product Description, Intended use, Application

The sterile set contains separately packed gauze balls made of dressing gauze complying with standard DIN EN 14079 type 20. They have a diameter of 3.0 cm.

The set can be used for disinfection when preparing the catheterization of the urine bladder. It is also suited for the disinfectant preparation and final treatment of the vascular access in the case of dialysis. The stable package has a perforation so that the chambers can be separately used according to requirements.

Composition

Cotton, silicone ring

Contraindications

The product should not be used in the case of a known allergy against the material.

Incident reporting

According to MDR (EU) 2017/745, if serious incidents occur in relation to the device, they must be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.

Normative and Regulative Requirements, Common Standards

Medical Device according to MDD 93/42/EEC, MDR (EU) 2017/745.

The product does not contain dangerous toxic substances according to REACH.

The gauze complies with the requirements of DIN EN 14079 type 20.

Sterilization complies with DIN EN 11135.

Packaging

Primary packaging: rigid film/paper packaging

Secondary packaging: folding box made of cellulose

Tertiary packaging: Carton made of cellulose

Symbols used in labelling

Explanations at www.nobamed.com



Marking on all packaging levels with CE and according to DIN EN ISO 15223-1- and ISO 20417.

Storage and Transport

Dry and dust free.

Sterile device

Before using a sterile product, visually inspect the packaging to ensure that it is intact.

Single use device

Reusing a single use medical device can lead to microbiological danger. Reprocessing for reuse can decrease the product's performance significantly.

Disposal

According to locally applicable legal regulations and standards of infection prophylaxis.