

# **NOBAMED Paul Danz AG**

# NOBACULAR-steril®

#### **REF 873150**

# Product Description, Intended use, Application

The oval, sterile eye ring pad, **size 88 x 95 mm**, with a punched oval opening is used after operations when no pressure should be exerted on the eye. The pad is also suited as padding with a watch-glass bandage. NOBACULAR®-sterile is for single use only and individually packed.

#### Composition

Polyester, viscose, rayon

#### **Contraindications**

The product should not be used in the case of a known allergy against the material.

#### Incident reporting

According to MDR (EU) 2017/745, if serious incidents occur in relation to the device, they must be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.

Normative and Regulative Requirements, Common Standards
Medical Device according to MDD 93/42/EEC, MDR (EU) 2017/745.

Sterilization of the product complies with DIN EN 11135.

The product does not contain dangerous toxic substances according to REACH.

#### **Packaging**

Primary packaging:

paper-film packaging

Secondary packaging:

folding box made of cellulose

Tertiary packaging:

carton made of cellulose

## Symbols used in labelling

Explanation at www.nobamed.com









**(€** 0123

Marking on all packaging levels with CE and according to DIN EN ISO 15223-1- and DIN EN ISO 20417

### **Storage and Transport**

Dry and dustfree

#### Sterile device

Before using a sterile product, visually inspect the packaging to ensure that it is intact.

### Single use device

Reusing a single use medical device can lead to microbiological danger. Reprocessing for reuse can decrease the product's performance significantly.

### Disposal

According to locally applicable legal regulations and standards of infection prophylaxis.

Product Data Sheet: November 10, 2022 [REV 4] replaces: September 24, 2015 [REV 3]