

# **NOBAMED Paul Danz AG**

# NOBADRAINAL®-steril

#### **REF 874355**

# **Product Description, Intended Use, Application**

NOBADRAINAL®-steril is a swab with Y-cut, size 4 cm x 4 cm and provided with a punched hole. It can therefore be used for hole diameter of 5-10 mm and for probes, catheters, drainages and extensions with an outside diameter of 9–30 Charrière (Ch). The sterile, aluminium-coated, and thereby non–adhesive drainage swab is used for the outlet of probes, catheters, drainages and extensions. NOBADRAINAL®-sterile is for single use only and individually packed.

# Composition

Polyethylen, aluminium, polyester, rayon

# **Contraindications**

The product should not be used in the case of a known allergy against the material.

#### Incident reporting

According to MDR (EU) 2017/745, if serious incidents occur in relation to the device, they must be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.

# Normative and Regulative Requirements, Common Standards

Medical Device according to MDD 93/42/EEC, MDR (EU) 2017/745.

Sterilization of the product complies with DIN EN 11135.

The product does not contain dangerous toxic substances according to REACH.

#### **Packaging**

Primary packaging: paper-paper

packaging

Secondary packaging: folding box

made of cellulose

Tertiary packaging: carton made of cellulose

# Symbols used in labelling

Explanation at www.nobamed.com









Marking on all packaging levels with CE and according to DIN EN ISO 15223-1- and DIN EN ISO 20417

### **Storage and Transport**

Dry and dustfree

### Sterile device

Before using a sterile product, visually inspect the packaging to ensure that it is intact.

### Single use device

Reusing a single use medical device can lead to microbiological danger. Reprocessing for reuse can decrease the product's performance significantly.

### Disposal

According to locally applicable legal regulations and standards of infection prophylaxis

Product Data Sheet: April 06, 2022 [REV 5] replaces: September 24, 2015 [REV 4]